Video Touch 4K and MDR Video Touch Release Notes

PG11025 Rev012

Date: October 2020

Versions: Video Touch 4K 3.0 and MDR Video Touch 1.32

Copyright

Information in this document is subject to change without notice. The software described in this document is furnished only under a separate license agreement and may be used or copied only according to the terms of such agreement. It is against the law to copy the software except as specifically allowed in the license agreement. This document or accompanying materials contains certain information which is confidential information of Hyland Software, Inc. and its affiliates, and which is subject to the confidentiality provisions agreed to by you.

All data, names, and formats used in this document's examples are fictitious unless noted otherwise. Complying with all applicable copyright laws is the responsibility of the user. Without limiting the rights under copyright law, no part of this document may be reproduced, stored in or introduced into a retrieval system, or transmitted in any form or by any means (electronic, mechanical, photocopying, recording, or otherwise), or for any purpose, without the express written permission of Hyland Software, Inc. or one of its affiliates.

Hyland®, Hyland Software®, Hyland Healthcare, and Hyland product names, are registered and/or unregistered trademarks of Hyland Software, Inc. and its affiliates in the United States and other countries. All other trademarks, service marks, trade names and products of other companies are the property of their respective owners.

© 2020 Hyland Software, Inc. and its affiliates. All rights reserved.

Regulatory and compliance

Australian Sponsor:
Emergo Australia
Level 20 Tower II
Darling Park 201 Sussex Street
Sydney, NSW 2000 Australia

EU Representative:

EC REP

EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands

MDR Video Touch
Anvisa n°: 80117580277
Responsável Technico: Luiz Levy Cruz Martins
CRF-SP: 42415
IMPORTADOR:
Emergo Brazil Import Importação e Distribuição de Produtos Médicos Hospitalares Ltda.
Avenida Francisco Matarazzo, 1.752, Salas 502/503, Água Branca, São Paulo-SP,
CEP - 05001-200
CNPJ: 04.967.408/0001-98
e-mail: brazilvigilance@ul.com

Manufacturer
Hyland LLC
12919 Earhart Ave
Auburn, CA 95602 USA

Support
Telephone: 1-844-535-1404
Web: https://community.hyland.com

General warnings and precautions

To ensure patient safety and adhere to HIPAA Privacy Rule involving patient health information, the Healthcare provider shall assign a unique Accession Number and Medical Record Number (MRN) to identify individual patient's health record affiliated with the patient to prevent erroneous medical data and duplicate identification with other patients.

Symbols glossary

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

3082 - Manufacturer: Indicates the medical device manufacturer.

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

Authorized representative in European Community: Indicates the authorized representative in the European union. Symbol is accompanied by the name and address of the authorized representative adjacent to the symbol.

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

2497 - Date of Manufacture: Indicates the date when the medical device was manufactured.

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

2493 - Catalog number: Indicates the manufacturer's catalog number so the medical device can be identified.

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

2498 - Serial number: Indicates the manufacturer's serial number so that a specific medical device can be identified.

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

1641 - Consult instructions for use: Indicates the need for the user to consult the instructions for use.

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

0434A - Caution: Indicates the need for the user to consult the instructions for use for important information such as warnings and cautions.

ISO - 15223-1 - Medical devices - Symbols to be used with medical device labels, labeling, and information to be supplied.

2610 - Patient number: Indicates a unique number associated with an individual patient.

Medical Devices Directive 93/42/EEC

CE Marking: Indicates that a product complies with applicable European Union regulations.

Quantity: Indicates the number of units.

Directive 2012/19/EU on waste electrical and electronic equipment (WEEE).

Table of Contents

Copyright ... **2**
Regulatory and compliance ... **3**
General warnings and precautions ... **3**
Symbols glossary ... **4**
What are the Video Touch 4K and MDR Video Touch? ... **6**
MDR Video Touch ... **6**
MDR Video Touch 1.32 (August 2020) ... 6
Updates ... *6*
MDR Video Touch 1.31 (December 2020) ... 6
Features ... *6*
Updates ... *6*
MDR Video Touch 1.3 (July 2019) ... 6
Features ... *6*
Updates ... *7*
MDR Video Touch 1.21 (December 2018) ... 7
Features ... *7*
Updates ... *7*
MDR Video Touch 1.2 (November 2018) ... 7
Features ... *7*
Updates ... *8*
MDR Video Touch 1.1.0 (August 2017) ... 8
Features ... *8*
Updates ... *8*
Video Touch 4K ... **8**
MDR Video Touch 3.0 (October 2020) ... 8
Features ... *8*
Updates ... *8*
Video Touch 4K 2.0.24 (January 2020) ... 9
Features ... *9*
Updates ... *9*
MDR Video Touch 2.0 (December 2019) ... 9
Features ... *9*
Updates ... *9*

What are the Video Touch 4K and MDR Video Touch?

The Video Touch 4K and MDR Video Touch are the hardware devices that captures and sends high-resolution images and video from a wide range of modalities to PACS/VNA.They provide image and video capture and transfer at the point of care without the need for a workstation.

The Video Touch 4K and MDR Video Touch software are functionally the same, even though they have different hardware. When Video Touch 4K is specified, it may be assumed that this also applies to MDR Video Touch, unless it specifies that it apply to only one product.

The Video Touch 4K and MDR Video Touch have a built-in server that handles security, configuration, and the flow of data between the modalities and destination DICOM systems. However, the Video Touch 4K and MDR Video Touch can also connect to an external instance of PACSgear Enterprise Server to handle these functions.

To configure PACSgear Enterprise Server settings, administrators can use a browser-based interface called Core Server Console (CS Console). You can launch CS Console from a PC or within the Video Touch 4K and MDR Video Touch.

MDR Video Touch

MDR Video Touch 1.32 (August 2020)

Updates

- Resolved an issue where the timestamp was incorrectly written with Epoch time in GMT and not the local time zone.

MDR Video Touch 1.31 (December 2020)

Features

- Added auto resend timer setting for failed jobs in serversettings.xml configuration file.

Updates

- Resolved an issue where pending jobs would get stuck.
- Resolved an issue where users were unable to resend failed jobs manually.

MDR Video Touch 1.3 (July 2019)

Features

- Added an icon to the Login and Patient Demographic screens to indicate the current network connection status.
- Added the MDR Link option with the remote PACSGear Enterprise Server 4.2 or later.
- The Login password now displays as asterisks when an external keyboard is plugged in.
- Added a Logoff confirmation dialog box.
- Added Tooltips to the new network icon.

2020 Hyland LLC

- Added log file entry when the network status changes and accessing the PACSGear Enterprise Server fails, such as login, renew session, testing connections failures.

Updates

- Resolved an issue that prevented users from using the dot character '.' in their login password.
- Resolved an issue that prevented users from playing recorded video after a power cycle.
- Resolved an issue that caused the system to display an unhandled exception error when a patient's gender was not an abbreviated string, such as MALE or FEMALE.
- Resolved an issue where a 3-characters modality string, such as PDF, was incorrectly cut off to a 2-characters string.
- Resolved an issue where the user login failed if the Department lookup sources names summary string was more than 255 characters.
- Updated the default number of days to purge data folders from 7 days to 30 days
- Updated the purge data process to only delete folders when they are empty.
- Resolved an issue with sending jobs in offline mode when unit is turned on.

MDR Video Touch 1.21 (December 2018)

Features

- Worklists are now available offline.
- The system now displays a warning alert when the user presses the Start Exam button without entering the patient's demographics.

Updates

- Resolved an issue on the Image Edit screen that prevented the user from canceling exams containing images.
- Resolved an issue where the MDR-VT box incorrectly rebooted even when there were no profiles to process.

MDR Video Touch 1.2 (November 2018)

Features

- You can now save images in bitmap (BMP) format.
- Added support for international languages.
- When a Send job from the client to the Service fails, the system now resends the metadata for the job on retry.
- Added a Multiple Destinations menu when sending a study to the destination archive.

Updates

- Resolved an issue that caused the system sound to unmute after a reboot.
- Resolved an issue that caused the system to display a new Capture screen after the user pressed the Enter or space bar keys.
- Resolved an issue that caused an unhandled exception if a lookup source contained an incorrect date/time format.

MDR Video Touch 1.1.0 (August 2017)

Features

- MDR-VT now supports capturing video in AVI format.

Updates

- The caret '^' character is now supported in the MRN field.
- The Numeric VKB now has a '-' key.
- Fixed an issue where UI distortion occurs on the Record form
- The Referring Physician name tag (0008,0090) value now properly passes to the PACS archive
- Fixed an unhandled exception when tapping Alt key on the virtual keyboard

Video Touch 4K

MDR Video Touch 3.0 (October 2020)

Features

- MDR is now a 64-bit application.
- Added a Domain list to the Login screen.
- Added image preview to the Record screen.
- Added the Allow Image setting in PACSgear Enterprise Server. This setting determines whether the Capture Image button is enabled.
- Added an Encounter number when sending data to non-DICOM destinations.
- Added Filmstrip overlay for video thumbnails.

Updates

- Resolved an issue where jobs with UNKNOWN status were incorrectly deleted.
- Resolved an issue the system would resend the same Failed job twice.
- Resolved an issue where the Default department would not display at the top of the Departments list.
- When clicking Send on the Edit screen, the system now displays a confirmation dialog box before sending.

Video Touch 4K 2.0.24 (January 2020)

Features

- Added a new auto resend timer setting for failed jobs in the serversettings.xml configuration file.
- Added support for Windows 10 Enterprise LTSC.
- Added support for HEVC/H.265 video recording.
- Added support, with PACSgear Enterprise Server 5.0 or later, for HEVC/H.265 video transfer.

Updates

- Resolved an issue, with remote PACSgear Enterprise Server 4.2.1 or later, where studies would not query by modality from ImageLink worklist.
- Resolved an issue where Internet Time would not synchronize.
- Resolved an issue where pending jobs would get stuck in a running state.
- Resolved an issue where failed jobs could not be resent manually.

MDR Video Touch 2.0 (December 2019)

Features

- Added support for Windows 10 Enterprise LTSC.
- Added support for HEVC/H.265 video format.

Updates

- Resolved an issue, with remote PACSgear Enterprise Server 4.2.1 or later, where studies would not query by modality from ImageLink worklist.
- Resolved an issue where Internet Time would not synchronize.